

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.089 °C	-25.3 °C	-0.211 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	0.310 °C	0.2 °C	-0.110 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.194 °C	5.2 °C	0.006 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.107 °C	10.5 °C	0.393 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.0 °C	-0.249 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.163 °C	30.3 °C	0.137 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.641 °C	41.0 °C	0.359 °C	± 0.5 °C	± 4.0E-01 °C

