

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.089 °C        | -25.4 °C  | -0.311 °C     | ± 0.70 °C            | ± 4.0E-01 °C                |
| 0 °C              | 0.310 °C          | 0.2 °C    | -0.110 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 5.194 °C          | 4.9 °C    | -0.294 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.107 °C         | 10.0 °C   | -0.107 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.249 °C         | 25.5 °C   | 0.251 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.163 °C         | 30.6 °C   | 0.437 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.641 °C         | 40.9 °C   | 0.259 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |

