

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.242 °C	14.9 °C	-0.342 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.1 °C	-0.263 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	30.0 %HR	-0.366 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	70.0 %HR	-0.448 %HR	± 5.0 %HR	± 1.9E+00 %HR