

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.070 °C	29.06 °C	-0.010 °C	± 0.2 °C	± 4.0E-01 °C
30 %HR	29.824 %HR	25.24 %HR	-4.584 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.813 %HR	43.53 %HR	-6.283 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	73.904 %HR	61.94 %HR	-11.964 %HR	± 5.0 %HR	± 1.9E+00 %HR

Resultados (Results) Punto de Rocio

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	20.666 °C	18.63 °C	-2.04 °C	± 2.0 °C	± 4.0E-01 °C
10 °C	9.045 °C	6.94 °C	-2.11 °C	± 2.0 °C	± 4.0E-01 °C