

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.5	0.49922	0.780	± 0.016	± 4.3E-01 mm					
0 a 1.5	1.49922	0.780	± 0.8	± 4.8E-01 mm					
0 a 2	1.99890	1.100	± 1.2	± 4.8E-01 mm					
0 a 3	2.99916	0.840	± 1.4	± 4.8E-01 mm					
0 a 4	3.99912	0.880	± 1.8	± 4.8E-01 mm					
0 a 5	4.99908	0.920	± 2.2	± 4.8E-01 mm					
0 a 6	5.99934	0.660	± 2.6	± 3.6E-01 mm					
0 a 7	6.99934	0.660	± 3	± 3.6E-01 mm					
0 a 8	7.99928	0.720	± 3.4	± 3.6E-01 mm					
0 a 9	8.99936	0.640	± 3.8	± 3.6E-01 mm					
0 a 10	9.99956	0.440	± 4.2	± 3.6E-01 mm					
0 a 11	10.99951	0.490	± 4.6	± 3.6E-01 mm					
0 a 12	11.99953	0.470	± 5	± 3.6E-01 mm					
0 a 13	12.99942	0.580	± 5.4	± 3.6E-01 mm					
0 a 14	13.99974	0.260	± 5.8	± 3.6E-01 mm					
0 a 15	14.99957	0.430	± 6.2	± 3.6E-01 mm					
0 a 16	15.99958	0.420	± 6.6	± 3.6E-01 mm					
0 a 17	16.99976	0.240	± 7	± 3.6E-01 mm					
0 a 18	17.99955	0.450	± 7.4	± 3.6E-01 mm					
0 a 19	18.99968	0.320	± 7.8	± 3.6E-01 mm					
0 a 20	19.99990	0.100	± 8.2	± 3.6E-01 mm					