

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.5	0.49922	0.780		± 4.3E-01 mm					
0 a 1.5	1.49922	0.780		± 4.8E-01 mm					
0 a 2	1.99929	0.710		± 4.8E-01 mm					
0 a 3	2.99916	0.840		± 4.8E-01 mm					
0 a 4	3.99912	0.880		± 4.8E-01 mm					
0 a 5	4.99908	0.920		± 4.8E-01 mm					
0 a 6	5.99934	0.660		± 3.6E-01 mm					
0 a 7	6.99934	0.660		± 3.6E-01 mm					
0 a 8	7.99928	0.720		± 3.6E-01 mm					
0 a 9	8.99936	0.640		± 3.6E-01 mm					
0 a 10	9.99956	0.440		± 3.6E-01 mm					
0 a 11	10.99951	0.490		± 3.6E-01 mm					
0 a 12	11.99953	0.470		± 3.6E-01 mm					
0 a 13	12.99942	0.580		± 3.6E-01 mm					
0 a 14	13.99974	0.260		± 3.6E-01 mm					
0 a 15	14.99957	0.430		± 3.6E-01 mm					
0 a 16	15.99958	0.420		± 3.6E-01 mm					
0 a 17	16.99976	0.240		± 3.6E-01 mm					
0 a 18	17.99955	0.450		± 3.6E-01 mm					
0 a 19	18.99968	0.320		± 3.6E-01 mm					
0 a 20	19.99990	0.100		± 3.6E-01 mm					