

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
12 °C	12.614 °C	12.4 °C	-0.214 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	30.042 °C	29.7 °C	-0.342 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.296 °C	42.6 °C	0.304 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.758 %HR	54.6 %HR	1.842 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	73.893 %HR	75.5 %HR	1.607 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.636 %HR	94.5 %HR	1.864 %HR	± 3.0 %HR	± 1.9E+00 %HR