

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 12 °C                      | 12.614 °C                  | 13.0 °C            | 0.386 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 27 °C                      | 30.042 °C                  | 29.7 °C            | -0.342 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 42 °C                      | 42.296 °C                  | 41.9 °C            | -0.396 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 52 %HR                     | 52.758 %HR                 | 55.4 %HR           | 2.642 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 72 %HR                     | 73.893 %HR                 | 76.4 %HR           | 2.507 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| # 92 %HR                   | 92.636 %HR                 | 95.2 %HR           | 2.564 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |