

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.200 °C	0.1 °C	-0.100 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.174 °C	5.0 °C	-0.174 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	9.8 °C	-0.207 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.926 %HR	31.5 %HR	1.574 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.713 %HR	52.8 %HR	2.087 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.221 %HR	71.2 %HR	1.979 %HR	± 3.0 %HR	± 1.9E+00 %HR