

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.200 °C	0.3 °C	0.100 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.174 °C	4.9 °C	-0.274 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.2 °C	0.193 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.926 %HR	32.4 %HR	2.474 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.713 %HR	53.4 %HR	2.687 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.221 %HR	72.1 %HR	2.879 %HR	± 3.0 %HR	± 1.9E+00 %HR