

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
12 °C	12.614 °C	13.0 °C	0.386 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	30.042 °C	29.8 °C	-0.242 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.296 °C	42.7 °C	0.404 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.758 %HR	50.4 %HR	-2.358 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	73.893 %HR	71.2 %HR	-2.693 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.636 %HR	90.8 %HR	-1.836 %HR	± 3.0 %HR	± 1.9E+00 %HR