

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.121 °C	15.0 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.2 °C	0.138 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	50.263 %HR	50.4 %HR	0.137 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	70.1 %HR	-0.143 %HR	± 3.0 %HR	± 1.9E+00 %HR