

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.121 °C	14.8 °C	-0.321 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.0 °C	-0.062 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	50.263 %HR	51.4 %HR	1.137 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	72.6 %HR	2.357 %HR	± 5.0 %HR	± 1.9E+00 %HR