

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.121 °C	14.8 °C	-0.321 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.3 °C	0.238 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	50.263 %HR	50.9 %HR	0.637 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	72.6 %HR	2.357 %HR	± 5.0 %HR	± 1.9E+00 %HR