

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.363 °C	30.4 °C	0.037 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	30.6 %HR	0.439 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.396 %HR	50.2 %HR	-0.196 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	70.4 %HR	-0.048 %HR	± 3.0 %HR	± 1.9E+00 %HR