

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 30 °C             | 30.363 °C         | 30.4 °C   | 0.037 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.161 %HR        | 31.5 %HR  | 1.339 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.396 %HR        | 50.5 %HR  | 0.104 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.448 %HR        | 68.5 %HR  | -1.948 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |