

7/9/2025

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.363 °C	30.1 °C	-0.263 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	31.5 %HR	1.339 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.396 %HR	50.3 %HR	-0.096 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.3 %HR	0.852 %HR	± 3.0 %HR	± 1.9E+00 %HR