

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
60 RPM	60 RPM	0.0 RPM	± 0.13 RPM	±	5.8E-02 RPM
600 RPM	600 RPM	0.0 RPM	± 0.40 RPM	±	5.8E-02 RPM
6000 RPM	6000.2 RPM	0.2 RPM	± 3.1 RPM	±	5.8E-02 RPM
60000 RPM	60003 RPM	3 RPM	± 31 RPM	±	8.0E-01 RPM
99960 RPM	99963 RPM	3 RPM	± 51 RPM	±	8.4E-01 RPM