

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -25 °C	-25.260 °C	-24.9 °C	0.360 °C	± 0.5 °C	± 4.0E-01 °C
# -20 °C	-20.286 °C	-20.0 °C	0.286 °C	± 0.5 °C	± 4.0E-01 °C
# -10 °C	-10.178 °C	-9.9 °C	0.278 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.171 °C	-0.1 °C	0.071 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.753 °C	5.0 °C	0.247 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.777 °C	10.2 °C	0.423 °C	± 0.5 °C	± 4.0E-01 °C