

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
0.10 mm	0.10 mm	0 µm	± 30 µm	±	5.8E+00 µm
0.20 mm	0.20 mm	0 µm	± 30 µm	±	5.8E+00 µm
0.30 mm	0.30 mm	0 µm	± 30 µm	±	5.8E+00 µm
0.40 mm	0.40 mm	0 µm	± 30 µm	±	5.8E+00 µm
0.50 mm	0.50 mm	0 µm	± 30 µm	±	5.8E+00 µm
0.60 mm	0.60 mm	0 µm	± 30 µm	±	5.8E+00 µm
0.70 mm	0.70 mm	0 µm	± 30 µm	±	5.8E+00 µm
0.80 mm	0.81 mm	10 µm	± 30 µm	±	5.8E+00 µm
1.00 mm	1.01 mm	10 µm	± 30 µm	±	5.8E+00 µm