

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
26.00 mm	26.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
26.25 mm	26.25 mm	0 μm	± 4 μm	±	5.8E+00 μm
26.50 mm	26.50 mm	0 μm	± 4 μm	±	5.8E+00 μm
27.00 mm	27.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
28.00 mm	28.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
29.00 mm	29.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
30.00 mm	30.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
35.00 mm	35.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
40.00 mm	40.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
45.00 mm	45.00 mm	0 μm	± 4 μm	±	5.8E+00 μm
50.00 mm	50.00 mm	0 μm	± 4 μm	±	5.8E+00 μm