

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
51.00 mm	51.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
51.25 mm	51.25 mm	0 μm	± 5 μm	±	5.8E+00 μm
51.50 mm	51.50 mm	0 μm	± 5 μm	±	5.8E+00 μm
52.00 mm	52.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
53.00 mm	53.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
54.00 mm	54.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
55.00 mm	55.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
60.00 mm	60.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
65.00 mm	65.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
70.00 mm	70.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
75.00 mm	75.00 mm	0 μm	± 5 μm	±	5.8E+00 μm