

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
76.00 mm	76.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
76.25 mm	76.25 mm	0 μm	± 5 μm	±	5.8E+00 μm
76.50 mm	76.50 mm	0 μm	± 5 μm	±	5.8E+00 μm
77.00 mm	77.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
78.00 mm	78.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
79.00 mm	79.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
80.00 mm	80.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
85.00 mm	85.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
90.00 mm	90.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
95.00 mm	95.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
100.00 mm	100.00 mm	0 μm	± 5 μm	±	5.8E+00 μm