

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
101.00 mm	101.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
101.25 mm	101.25 mm	0 μm	± 6 μm	±	5.8E+00 μm
101.50 mm	101.50 mm	0 μm	± 6 μm	±	5.8E+00 μm
102.00 mm	102.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
103.00 mm	103.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
104.00 mm	104.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
105.00 mm	105.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
110.00 mm	110.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
115.00 mm	115.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
120.00 mm	120.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
125.00 mm	125.00 mm	0 μm	± 6 μm	±	5.8E+00 μm