

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
126.00 mm	126.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
126.25 mm	126.25 mm	0 μm	± 6 μm	±	5.8E+00 μm
126.50 mm	126.50 mm	0 μm	± 6 μm	±	5.8E+00 μm
127.00 mm	127.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
128.00 mm	128.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
129.00 mm	129.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
130.00 mm	130.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
135.00 mm	135.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
140.00 mm	140.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
145.00 mm	145.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
150.00 mm	150.00 mm	0 μm	± 6 μm	±	5.8E+00 μm