

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 30 °C             | 30.062 °C         | 30.5 °C   | 0.438 °C      | ± 1.0 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.263 %HR        | 32.0 %HR  | 1.737 %HR     | ± 5.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.642 %HR        | 49.0 %HR  | -1.642 %HR    | ± 5.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.448 %HR        | 70.0 %HR  | -0.448 %HR    | ± 5.0 %HR            | ± 1.9E+00 %HR               |