

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.9457 µl	2 µl	0.054 µl	± 0.1 µl	±	5.3E-02 µl
5.0523 µl	5 µl	-0.052 µl	± 0.1 µl	±	5.1E-02 µl
9.909 µl	10 µl	0.091 µl	± 0.1 µl	±	5.2E-02 µl