

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 1 mm              | 1.00 mm   | 0 µm          | ± 20 µm           | ±                           | 5.8E+00 µm |
| 1.25 mm           | 1.25 mm   | 0 µm          | ± 20 µm           | ±                           | 5.8E+00 µm |
| 1.5 mm            | 1.50 mm   | 0 µm          | ± 20 µm           | ±                           | 5.8E+00 µm |
| 2 mm              | 2.00 mm   | 0 µm          | ± 20 µm           | ±                           | 5.8E+00 µm |
| 3 mm              | 3.00 mm   | 0 µm          | ± 20 µm           | ±                           | 5.8E+00 µm |
| 4 mm              | 4.01 mm   | 10 µm         | ± 20 µm           | ±                           | 5.8E+00 µm |
| 5 mm              | 5.01 mm   | 10 µm         | ± 20 µm           | ±                           | 5.8E+00 µm |
| 8 mm              | 7.99 mm   | -10 µm        | ± 20 µm           | ±                           | 5.8E+00 µm |
| 10 mm             | 9.99 mm   | -10 µm        | ± 20 µm           | ±                           | 5.8E+00 µm |