

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.00 mm	1.02 mm	20 µm	± 30 µm	±	5.8E+00 µm
1.50 mm	1.52 mm	20 µm	± 30 µm	±	5.8E+00 µm
2.00 mm	2.03 mm	30 µm	± 30 µm	±	5.8E+00 µm
4.00 mm	4.02 mm	20 µm	± 30 µm	±	5.8E+00 µm
8.00 mm	8.02 mm	20 µm	± 30 µm	±	5.8E+00 µm
10.00 mm	10.02 mm	20 µm	± 30 µm	±	5.8E+00 µm
15.00 mm	15.03 mm	30 µm	± 30 µm	±	5.8E+00 µm
20.00 mm	20.02 mm	20 µm	± 30 µm	±	5.8E+00 µm
25.00 mm	25.03 mm	30 µm	± 30 µm	±	5.8E+00 µm