

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.3 μm	26.0 μm	2.70 μm	± 3 μm	±	1.9E+02 μm
51.6 μm	55.0 μm	3.40 μm	± 3 μm	±	1.9E+02 μm
124.6 μm	129.0 μm	4.40 μm	± 3 μm	±	1.9E+02 μm
250.3 μm	255.0 μm	4.70 μm	± 3 μm	±	1.9E+02 μm
486.00 μm	490.0 μm	4.00 μm	± 3 μm	±	1.9E+02 μm
963.00 μm	969.0 μm	6.00 μm	± 3 μm	±	1.9E+02 μm