

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.2	0.19974	0.260	± 0.6	±	4.3E-01 mm						
0 a 0.5	0.49960	0.400	± 0.8	±	4.3E-01 mm						
0 a 1	0.99958	0.420	± 1	±	4.3E-01 mm						
0 a 2	1.99948	0.520	± 1.4	±	4.8E-01 mm						
0 a 3	2.99946	0.540	± 1.8	±	4.8E-01 mm						
0 a 4	3.99934	0.660	± 2.2	±	4.8E-01 mm						
0 a 5	4.99918	0.820	± 2.6	±	4.8E-01 mm						
0 a 6	5.99903	0.970	± 3	±	3.6E-01 mm						
0 a 7	6.99895	1.050	± 3.4	±	3.6E-01 mm						
0 a 8	7.99870	1.300	± 3.8	±	3.6E-01 mm						