

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
505 RPM	500 RPM	-5 RPM	± 20 RPM	±	5.8E+01 RPM
1011 RPM	1000 RPM	-11 RPM	± 20 RPM	±	5.8E+01 RPM
1518 RPM	1500 RPM	-18 RPM	± 20 RPM	±	5.8E+01 RPM
2024 RPM	2000 RPM	-24 RPM	± 20 RPM	±	5.8E+01 RPM
3032 RPM	3000 RPM	-32 RPM	± 20 RPM	±	5.8E+01 RPM