

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	1 mm	1.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	5 mm	5.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	10 mm	10.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	15 mm	15.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	20 mm	20.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	25 mm	25.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	30 mm	30.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	35 mm	35.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	40 mm	40.0 mm	0 µm	± 50 µm	± 5.8E+01 µm