

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10 RPM	10 RPM	0 RPM	± 3 RPM	±	5.8E-01 RPM
20.1 RPM	20 RPM	-0.1 RPM	± 3 RPM	±	5.8E-01 RPM
50.2 RPM	50 RPM	-0.2 RPM	± 3 RPM	±	5.8E-01 RPM
99.9 RPM	100 RPM	0.1 RPM	± 3 RPM	±	5.8E+01 RPM