

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
29.5 RPM	28 RPM	-1.5 RPM	± 5 RPM	±	1.6E+01 RPM
85.1 RPM	84 RPM	-1.1 RPM	± 5 RPM	±	1.6E+01 RPM
138.1 RPM	140 RPM	1.9 RPM	± 5 RPM	±	1.6E+01 RPM
197.9 RPM	196 RPM	-1.9 RPM	± 5 RPM	±	5.8E+01 RPM
277.6 RPM	280 RPM	2.4 RPM	± 5 RPM	±	5.8E+01 RPM