

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.021 °C	15.0 °C	-0.021 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.149 °C	25.0 °C	-0.149 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.318 °C	49.0 °C	-1.318 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.151 %HR	16.0 %HR	-9.151 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.642 %HR	40.0 %HR	-10.642 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.356 %HR	65.0 %HR	-10.356 %HR	± 5.0 %HR	± 1.9E+00 %HR