

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.021 °C	14.7 °C	-0.321 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.149 °C	25.2 °C	0.051 °C	± 0.5 °C	± 4.0E-01 °C
50 °C	50.318 °C	50.0 °C	-0.318 °C	± 0.5 °C	± 1.1E+00 °C
25 %HR	25.151 %HR	27.3 %HR	2.149 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.642 %HR	52.5 %HR	1.858 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.356 %HR	73.2 %HR	-2.156 %HR	± 3.0 %HR	± 1.9E+00 %HR