

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 15.021 °C                  | 15.3 °C            | 0.279 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.149 °C                  | 25.2 °C            | 0.051 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 50 °C                      | 50.318 °C                  | 50.8 °C            | 0.482 °C               | ± 0.5 °C                      | ± 1.1E+00 °C                         |
| 25 %HR                     | 25.151 %HR                 | 25.2 %HR           | 0.049 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.642 %HR                 | 50.7 %HR           | 0.058 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 75 %HR                     | 75.356 %HR                 | 75.2 %HR           | -0.156 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |