

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 12 °C             | 12.013 °C         | 12.0 °C   | -0.013 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 27 °C             | 27.024 °C         | 26.8 °C   | -0.224 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 42 °C             | 42.597 °C         | 42.6 °C   | 0.003 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 52 %HR            | 52.308 %HR        | 52.2 %HR  | -0.108 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 72 %HR            | 72.084 %HR        | 71.6 %HR  | -0.484 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| # 92 %HR          | 92.073 %HR        | 92.0 %HR  | -0.073 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |