

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 12 °C                      | 12.013 °C                  | 12.4 °C            | 0.387 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 27 °C                      | 27.024 °C                  | 27.0 °C            | -0.024 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 42 °C                      | 42.597 °C                  | 42.8 °C            | 0.203 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 52 %HR                     | 52.308 %HR                 | 53.6 %HR           | 1.292 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 72 %HR                     | 72.084 %HR                 | 71.8 %HR           | -0.284 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| # 92 %HR                   | 92.073 %HR                 | 92.5 %HR           | 0.427 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |