

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.013 °C	12.4 °C	0.387 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.024 °C	26.9 °C	-0.124 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.597 °C	42.9 °C	0.303 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.308 %HR	54.4 %HR	2.092 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.084 %HR	74.9 %HR	2.816 %HR	± 3.0 %HR	± 1.9E+00 %HR
# 92 %HR	92.073 %HR	94.5 %HR	2.427 %HR	± 3.0 %HR	± 1.9E+00 %HR