

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 0 °C                       | 0.130 °C                   | 0.5 °C             | 0.370 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 5 °C                       | 5.194 °C                   | 5.4 °C             | 0.206 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 10 °C                      | 10.158 °C                  | 10.4 °C            | 0.242 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 30.314 %HR                 | 32.8 %HR           | 2.486 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.263 %HR                 | 52.5 %HR           | 2.237 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.591 %HR                 | 72.2 %HR           | 1.609 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |