

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.130 °C	0.4 °C	0.270 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.194 °C	5.3 °C	0.106 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.158 °C	10.3 °C	0.142 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.314 %HR	39.2 %HR	8.886 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.263 %HR	59.4 %HR	9.137 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.591 %HR	79.1 %HR	8.509 %HR	± 3.0 %HR	± 1.9E+00 %HR