

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 0.05	0.04975	0.250	± 0.6	± 4.3E-01 mm					
0 a 0.1	0.09974	0.260	± 0.6	± 4.3E-01 mm					
0 a 0.15	0.14974	0.260	± 0.6	± 4.3E-01 mm					
0 a 0.5	0.49966	0.340	± 0.8	± 4.3E-01 mm					
0 a 1	0.99963	0.370	± 1	± 4.3E-01 mm					
0 a 2	1.99868	1.320	± 1.4	± 4.8E-01 mm					
0 a 3	2.99948	0.520	± 1.8	± 4.8E-01 mm					
0 a 4	3.99937	0.630	± 2.2	± 4.8E-01 mm					
0 a 5	4.99924	0.760	± 2.6	± 4.8E-01 mm					
0 a 6	5.99910	0.900	± 3	± 3.6E-01 mm					
0 a 7	6.99982	1.120	± 3.4	± 3.6E-01 mm					
0 a 8	7.99972	1.410	± 3.8	± 3.6E-01 mm					
0 a 9	8.99976	1.590	± 4.2	± 3.6E-01 mm					
0 a 10	9.99979	0.213	± 4.6	± 3.6E-01 mm					
0 a 11	10.99978	0.221	± 5	± 3.6E-01 mm					
0 a 12	11.99979	0.207	± 5.4	± 3.6E-01 mm					
0 a 13	12.99979	0.215	± 5.8	± 3.6E-01 mm					
0 a 14	13.99980	0.199	± 6.2	± 3.6E-04 mm					
0 a 15	14.99971	0.287	± 6.6	± 3.6E-04 mm					
0 a 16	15.99968	0.318	± 7	± 3.6E-04 mm					
0 a 17	16.99976	0.245	± 7.4	± 3.6E-04 mm					
0 a 18	17.99974	0.263	± 7.8	± 3.6E-04 mm					
0 a 19	18.99975	0.251	± 8.2	± 3.6E-04 mm					
0 a 20	19.99980	0.201	± 8.6	± 3.6E-04 mm					