

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.130 °C	0.5 °C	0.370 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.208 °C	10.4 °C	0.192 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.436 °C	20.8 °C	0.364 °C	± 3.0 °C	± 4.0E-01 °C
40 °C	40.291 °C	40.5 °C	0.209 °C	± 3.0 °C	± 4.0E-01 °C