

Certificado No. (Certificate #):58987.1Equipo (Instrument):Medidor de Espesores

Fecha de Calibración (Calibration Date):2025-07-28Marca (Brand):Defelsko

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.0000 μm	23.2500 μm	0.2500 μm		±	1.0E+00 μm
51.0000 μm	50.5000 μm	-0.5000 μm		±	1.4E+00 μm
75.0000 μm	74.0000 μm	-1.0000 μm		±	8.7E-01 μm
131.0000 μm	130.0000 μm	-1.0000 μm		±	8.7E-01 μm
182.0000 μm	180.0000 μm	-2.0000 μm		±	8.7E-01 μm
260.0000 μm	258.0000 μm	-2.0000 μm		±	8.7E-01 μm
387.0000 μm	384.0000 μm	-3.0000 μm		±	8.7E-01 μm
504.0000 μm	500.0000 μm	-4.0000 μm		±	8.8E-01 μm
761.0000 μm	756.0000 μm	-5.0000 μm		±	8.8E-01 μm