

Certificado No. (Certificate #):		58987.1		Equipo (Instrument):		Medidor de Espesores			
Fecha de Calibración (Calibration Date):		2025-07-28		Marca (Brand):		Defelsko			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
23.0000 μm		23 μm		0 μm				±	1.0E+00 μm
51.0000 μm		50 μm		-1 μm				±	1.4E+00 μm
75.0000 μm		74 μm		-1 μm				±	8.7E-01 μm
131.0000 μm		130 μm		-1 μm				±	8.7E-01 μm
182.0000 μm		180 μm		-2 μm				±	8.7E-01 μm
260.0000 μm		258 μm		-2 μm				±	8.7E-01 μm
387.0000 μm		384 μm		-3 μm				±	8.7E-01 μm
504.0000 μm		500 μm		-4 μm				±	8.8E-01 μm
761.0000 μm		756 μm		-5 μm				±	8.8E-01 μm