

Certificado No. (Certificate #):58987.1Equipo (Instrument):Medidor de Espesores

Fecha de Calibración (Calibration Date):2025-07-28Marca (Brand):Defelsko

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23 μm	23 μm	0 μm		±	1.0E+00 μm
51 μm	50 μm	-1 μm		±	1.4E+00 μm
75 μm	74 μm	-1 μm		±	8.7E-01 μm
131 μm	130 μm	-1 μm		±	8.7E-01 μm
182 μm	180 μm	-2 μm		±	8.7E-01 μm
260 μm	258 μm	-2 μm		±	8.7E-01 μm
387 μm	384 μm	-3 μm		±	8.7E-01 μm
504 μm	500 μm	-4 μm		±	8.8E-01 μm
761 μm	756 μm	-5 μm		±	8.8E-01 μm