

Certificado No. (Certificate #):		58987.2		Equipo (Instrument):		Medidor de Espesores			
Fecha de Calibración (Calibration Date):		2025-07-28		Marca (Brand):		Defelsko			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
100.0000 μm		96 μm		-4 μm				±	9.3E-01 μm
200.0000 μm		190 μm		-10 μm				±	1.9E+00 μm
300.0000 μm		289 μm		-11 μm				±	5.2E+00 μm
400.0000 μm		398 μm		-2 μm				±	4.7E+00 μm
410.0000 μm		408 μm		-2 μm				±	9.6E-01 μm
420.0000 μm		419 μm		-1 μm				±	7.7E-01 μm
450.0000 μm		453 μm		3 μm				±	9.6E-01 μm
480.0000 μm		487 μm		7 μm				±	2.4E+00 μm
500.0000 μm		513 μm		13 μm				±	1.4E+00 μm