

Certificado No. (Certificate #):58987.2Equipo (Instrument):Medidor de Espesores

Fecha de Calibración (Calibration Date):2025-07-28Marca (Brand):Defelsko

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
100 μm	96 μm	-4 μm		±	9.3E-01 μm
200 μm	190 μm	-10 μm		±	1.9E+00 μm
300 μm	289 μm	-11 μm		±	5.2E+00 μm
400 μm	398 μm	-2 μm		±	4.7E+00 μm
410 μm	408 μm	-2 μm		±	9.6E-01 μm
420 μm	419 μm	-1 μm		±	7.7E-01 μm
450 μm	453 μm	3 μm		±	9.6E-01 μm
480 μm	487 μm	7 μm		±	2.4E+00 μm
500 μm	513 μm	13 μm		±	1.4E+00 μm